



Immunoglobulin A

REF C9000150A 1x50ml
Turbidimetry

Quantitative determination of human immunoglobulin A (IgA) IVD

Store 2 - 8°C.

INTENDED USE

The IgA is a quantitative turbidimetric test for the measurement of IgA in human serum or plasma.

PRINCIPLE OF THE METHOD

Anti-human IgA antibodies when mixed with samples containing IgA, form insoluble complexes. These complexes cause an absorbance change, dependent upon the IgA concentration of the patient sample, that can be quantified by comparison from a calibrator of known IgA concentration.

CLINICAL SIGNIFICANCE

IgA represents approximately 10 to 15% of total serum immunoglobulins. Its structure is monomeric, similar to the IgG molecule, but 10 to 15% of IgA in serum is polymeric, particularly IgA₂, which is more resistant to destruction by some pathogenic bacteria. Another more important form of IgA is called secretory IgA. It is found in tears, sweat, saliva, milk and gastrointestinal and bronchial secretions.

IgA is generally increased in skin, pulmonary, kidney infections, and hepatic cirrhosis. Increased monoclonal IgA concentrations may be found in multiple myeloma and other disturbances of plasmatic cells.

REAGENTS

Diluent (R1)	Tris buffer 20 mmol/L, PEG 8000, pH 8.3 Sodium azide 0.95 g/L...
Antibody (R2)	Goat serum, anti-human IgA, pH 7.5. Sodium azida 0.95 g/L.

CALIBRATION

The assay is calibrated to the Reference Material CRM 470/RPPHS (Institute of Reference of Materials and Measurements, IRMM). It must be used the SERUM PROTEINS CALIBRATOR (Cod.:905CAL) to calibrate the reagent. The reagent (both monoreagent and bireagent) should be recalibrated every month, when the controls are out of specifications, and when changing the reagent lot or the instrument settings.

PREPARATION

Reagents: Ready to use.

Calibration Curve: Prepare the following SERUM PROTEINS CALIBRATOR Calibrator dilutions in NaCl 9 g/L as diluent. Multiply the concentration of the IgA calibrator by the corresponding factor stated in table below to obtain the IgA concentration of each dilution.

Calibrator dilution	1	2	3	4	5	6
Calibrator (µL)	--	10	25	50	75	100
NaCl 9 g/L (µL)	100	90	75	50	25	-
Factor	0	0.1	0.25	0.5	0.75	1.0

STORAGE AND STABILITY

All the components of the kit are stable until the expiration date on the label when stored tightly closed at 2-8°C and contaminations are prevented during their use. Do not use reagents over the expiration date.

Reagent deterioration: The presence of particles and turbidity.

Do not freeze; frozen Antibody or Diluent could change the functionality of the test.

ADDITIONAL EQUIPMENT

- Thermostatic bath at 37°C.
- Spectrophotometer or photometer thermostatable at 37°C with a 600 nm filter (580 – 620 nm).

SAMPLES

Fresh serum or plasma. EDTA or heparin should be used as anticoagulant. Stable 7 days at 2-8°C or 3 months at -20°C.

The samples with presence of fibrin should be centrifuged.

Do not use highly hemolyzed or lipemic samples.

PROCEDURE

- Bring the reagents and the photometer (cuvette holder) to 37°C.
- Assay conditions:
 - Wavelength : 600
 - Temperature : 37 °C
 - Cuvette ligh path : 1cm
- Adjust the instrument to zero with distilled water.
- Pipette into a cuvette:

Reagent R1	800 µL
Sample or Calibrator	10 µL

- Mix and read the absorbance (A_1) after the sample addition.
- Immediately, pipette into de cuvette:

Reagent R2	200 µL
------------	--------

- Mix and read the absorbance (A_2) of calibrators and sample exactly 2 minutes after the R2 addition.

Gesam has instruction sheets for several automatic analyzers. Instructions for many of them are available on request.

CALCULATIONS

Calculate the absorbance difference ($A_2 - A_1$) of each point of the calibration curve and plot the values obtained against the IgA concentration of each calibrator dilution. IgA concentration in the sample is calculated by interpolation of its ($A_2 - A_1$) in the calibration curve.

QUALITY CONTROL

Control sera are recommended to monitor the performance of manual and automated assay procedures. Gesam SERUM PROTEINS CONTROL (Cod.:905CTL). Each laboratory should establish its own Quality Control scheme and corrective actions if controls do not meet the acceptable tolerances.

REFERENCE VALUES⁴

Between 70 - 400 mg/dL. Each laboratory should establish its own reference range.

PERFORMANCE CHARACTERISTICS

- Measurement range: Up to 600 mg/dL, under the described assay conditions. Samples with higher concentrations, should be diluted 1/5 in NaCl 9 g/L and re-tested again. The linearity limit and measurement range depends on the sample to reagent / ratio. It will be higher by decreasing the sample volume, although the sensitivity of the test will be proportionally decreased.
- Limit detection: Values less than 1 mg/dL give non-reproducible results.
- Prozone effect: No prozone effect was detected upon 2000 mg/dL
- Sensitivity: $\Delta 2.1$ mA. mg/dL at 71 mg/dL.
- Precision: The reagent has been tested for 20 days, using three levels of serum in a EP5-based study.
- The final concentration of the components is below the limits imposed by Regulation (EC) No. 1272/2008 - CLP (and subsequent amendments) and Directive 88/379/CEE and subsequent amendments to the classification-packaging and labeling of dangerous substances.

EP5	CV (%)		
	127.7 mg/dl	196.9 mg/dl	416.3 mg/dl
Total	8.2%	5.2%	3.5%
Within Run	1.7%	1.5%	1%
Between Run	2.2%	1.9%	2.4%
Between Day	7.7%	4.6%	2.3%

- Accuracy: Results obtained using this reagent (y) were compared to those obtained using an immunoturbidimetric method from Bayer. 46 samples ranging from 20 to 400 mg/dL of IgA were assayed. The correlation coefficient (r) was 0.97 and the regression equation $y = 1.16x - 12.2$.

The results of the performance characteristics depend on the used analyzer.

INTERFERENCES

Hemoglobin (10 g/L), bilirubin (20 mg/dL) and lipemia (5 g/L), do not interfere. Rheumatoid factors may interfere at 900 IU/mL. Other substances may interfere^{6,7}.

NOTES

- Clinical diagnosis should not be made on findings of a single test result, but should integrate both clinical and laboratory data.

BIBLIOGRAPHY

- Clinical Guide to Laboratory Tests, Edited by NW Tietz W B Saunders Co., Philadelphia, 483, 1983.
- Skoug Jonh W et al. Clin Chem 1988; 34/2: 309 - 315
- Pesce AJ and Kaplan, LA. Methods in Clinical Chemistry. The CV Mosby Company, St. Louis MO, 1987.
- Dati F et al. Eur J Clin Chem Clin Biochem 1996; 34: 517-520.
- Young DS. Effects of disease on clinical laboratory tests, 3th ed. AACC Pres, 1997
- Friedman and Young. Effects of disease on clinical laboratory tests, 3th ed. AACC Pres, 1997.

PACKAGING

Ref.: 9000150	Cont.	: 1 x 40 mL R1. Diluent : 1 x 10 mL R2. Antibody
---------------	-------	---



CE Mark (98/79 CE regulation)



in vitro medical device



Batch Code



Use by



Storage temperature limits



Read instruction for use



Gesam Production srl